



17400 Medina Road
 Suite 100
 Minneapolis, MN
 763-398-8300 tel
 763-398-8400 fax
 carefusion.com

March 24, 2010

Gene Turner
 Sales Support Manager
 LTV Series Ventilators
 CareFusion
 17400 Medina Road
 Suite 100
 Minneapolis, MN 55447

To whom it may concern:

This letter and attachment are for your information and records concerning requirements for personal medical assistive devices used onboard passenger aircraft, specifically LTV Series Ventilators manufactured by CareFusion. (These ventilators may also be labeled under *Pulmonetic Systems*, *Cardinal Health*, or *Viasys* brand names.)

In accordance with the US Department of Transportation (DOT) Final Rule, "Non-discrimination on the Basis of Disability in Air Travel" (73 FR 27614 Which updates Title 14 CRF Part 382), effective May 13, 2009, we offer our Certificate of Compliance. This certificate confirms that all LTV Series Ventilators, LTMII Monitors, and SprintPack lithium ion battery systems meet the RTCA/DO-160F Section 20, Category S and Section 21, Category M requirements.

The following products are in compliance with the regulations stated above as attested to on the attached certification.

LTV 1200 Ventilators	LTV 900 Ventilators
LTV 1150 Ventilators	LTV 800 Ventilators
LTV 1000 Ventilators	LTMII Graphics Monitors
LTV 950 Ventilators	SprintPack battery systems

As the FAA has not released any guidelines concerning the contents of the required labeling on the ventilator despite repeated requests for such, we offer this certificate to the airlines and our customers. It is available from our website and through our manufacturer's representatives and customer service agents. If you have any questions or concerns regarding LTV Series ventilators used by a passenger, please contact myself, or our customer service technical advisors at (763) 398.8300.

Sincerely,

Gene Turner
 Sales Support Manager



"Your Agent to Product Certifications and Licenses Throughout the World"

RTCA/DO-160F

TEST CERTIFICATION

Manufacturer: CareFusion

Models: LTV 800, LTV 900, LTV 950, LTV 1000, LTV 1150, LTV1200, LTMII, SprintPack

Test Date: February 17, 2010

Issue Date: March 8, 2010

Test Report #170210

RTCA/DO-160F Sections 20, 21 Certification

The LTV Ventilator System has passed full compliance with the requirements of RTCA/DO-160F Section 20, Category S, and Section 21, Category M. Testing was performed using a Battery Power Source for the Ventilator and LTMII. Such certification is required for use aboard all commercial airlines landing and/or taking off within the United States. This certification complies with the Air Carriers Act and the Department of Transportation's Nondiscrimination on the Basis of Disability in Air Travel policy.

Approved by:

Duane R. Bagdons
President
International Certification Services Inc.
1100 Falcon Ave.
Glencoe, Minnesota 55336
(320) 864-4444

Customer Representative:

Gregory Smith
CareFusion
Infusion Respiratory Services
17400 Medina Road Suite 100
Minneapolis, Minnesota 55447-1341